



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-16808

Page No

1 of 9

Validity

22/09/2025 to 21/09/2029

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Testing				
1	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Exposure Rate at table top	AERB/RF-MED/SC-3 (Rev. 2)
2	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	High Contrast Resolution	AERB/RF-MED/SC-3 (Rev. 2)
3	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Irradiation Time	AERB/RF-MED/SC-3 (Rev. 2)
4	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Linearity of mA / mAs loading stations	AERB/RF-MED/SC-3 (Rev. 2)
5	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Low Contrast Resolution	AERB/RF-MED/SC-3 (Rev. 2)
6	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Operating Potential	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-16808

Page No

2 of 9

Validity

22/09/2025 to 21/09/2029

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
7	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Radiation Leakage Level from X-ray Tube Housing	AERB/RF-MED/SC-3 (Rev. 2)
8	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Radiation Protection Survey	AERB/RF-MED/SC-3 (Rev. 2)
9	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Reproducibility of Radiation Output (CoV)	AERB/RF-MED/SC-3 (Rev. 2)
10	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM & O-ARM	Total Filtration	AERB/RF-MED/SC-3 (Rev. 2)
11	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Exposure Rate at table top	AERB/RF-MED/SC-3 (Rev. 2)
12	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	High Contrast Resolution	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-16808

Page No

3 of 9

Validity

22/09/2025 to 21/09/2029

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
13	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Irradiation Time	AERB/RF-MED/SC-3 (Rev. 2)
14	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Linearity of mA / mAs loading stations	AERB/RF-MED/SC-3 (Rev. 2)
15	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Low Contrast Resolution	AERB/RF-MED/SC-3 (Rev. 2)
16	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Operating Potential	AERB/RF-MED/SC-3 (Rev. 2)
17	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Radiation Leakage Level from X-ray Tube Housing	AERB/RF-MED/SC-3 (Rev. 2)
18	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Radiation Protection Survey	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-16808 **Page No** 4 of 9

Validity 22/09/2025 to 21/09/2029 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
19	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Reproducibility of Radiation Output (CoV)	AERB/RF-MED/SC-3 (Rev. 2)
20	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology (Cath Lab)	Total Filtration	AERB/RF-MED/SC-3 (Rev. 2)
21	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Central Beam Alignment	AERB/RF-MED/SC-3 (Rev. 2)
22	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Congruence of Radiation & Optical Field	AERB/RF-MED/SC-3 (Rev. 2)
23	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Effective Focal Spot Measurement	AERB/RF-MED/SC-3 (Rev. 2)
24	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Exposure Rate at table top	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-16808	Page No	5 of 9
Validity	22/09/2025 to 21/09/2029	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
25	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	High Contrast Sensitivity	AERB/RF-MED/SC-3 (Rev. 2)
26	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Irradiation Time	AERB/RF-MED/SC-3 (Rev. 2)
27	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Linearity of mA / mAs loading stations	AERB/RF-MED/SC-3 (Rev. 2)
28	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Low Contrast Sensitivity	AERB/RF-MED/SC-3 (Rev. 2)
29	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Operating Potential	AERB/RF-MED/SC-3 (Rev. 2)
30	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Radiation Leakage Level from X-ray Tube Housing	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-16808

Page No

6 of 9

Validity

22/09/2025 to 21/09/2029

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
31	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Radiation Protection Survey	AERB/RF-MED/SC-3 (Rev. 2)
32	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Reproducibility of Radiation Output (CoV)	AERB/RF-MED/SC-3 (Rev. 2)
33	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy equipment	Total Filtration	AERB/RF-MED/SC-3 (Rev. 2)
34	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Central Beam Alignment	AERB/RF-MED/SC-3 (Rev. 2)
35	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Congruence of Radiation & Optical Field	AERB/RF-MED/SC-3 (Rev. 2)
36	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Effective Focal Spot Measurement	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-16808	Page No	7 of 9
Validity	22/09/2025 to 21/09/2029	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
37	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Irradiation Time	AERB/RF-MED/SC-3 (Rev. 2)
38	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Linearity of mA / mAs loading stations	AERB/RF-MED/SC-3 (Rev. 2)
39	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Operating Potential	AERB/RF-MED/SC-3 (Rev. 2)
40	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Radiation Leakage Level from X-ray Collimator	AERB/RF-MED/SC-3 (Rev. 2)
41	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Radiation Leakage Level from X-ray Tube Housing	AERB/RF-MED/SC-3 (Rev. 2)
42	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Radiation Protection Survey	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-16808

Page No

8 of 9

Validity

22/09/2025 to 21/09/2029

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
43	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Reproducibility of Radiation Output (CoV)	AERB/RF-MED/SC-3 (Rev. 2)
44	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	X-Ray Machine (FIXED, PORTABLE, MOBILE)	Total Filtration	AERB/RF-MED/SC-3 (Rev. 2)
45	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Bone mineral densitometer(BMD)& DEXA or DXA scan	Radiation Leakage Level from X-ray Tube Housing	AERB/RF-MED/SC-3 (Rev. 2)
46	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Bone mineral densitometer(BMD)& DEXA or DXA scan	Radiation Protection Survey	AERB/RF-MED/SC-3 (Rev. 2)
47	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Irradiation Time	AERB/RF-MED/SC-3 (Rev. 2)
48	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	mA/mAs Linearity (CoL)	AERB/RF-MED/SC-3 (Rev. 2)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

K2AG DIAGNOSTIC QUALITY ASSURANCE LLP, 24, SHIV MARKET, AZAD NAGAR, HISAR, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-16808

Page No

9 of 9

Validity

22/09/2025 to 21/09/2029

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
49	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Radiation Leakage Level from X-ray Tube Housing	AERB/RF-MED/SC-3 (Rev. 2)
50	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Radiation Protection Survey	AERB/RF-MED/SC-3 (Rev. 2)
51	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Reproducibility of Radiation Output (CoV)	AERB/RF-MED/SC-3 (Rev. 2)
52	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Timer linearity (Coefficient of Linearity)	AERB/RF-MED/SC-3 (Rev. 2)
53	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Total Filtration	AERB/RF-MED/SC-3 (Rev. 2)
54	DIAGNOSTIC RADIOLOGY QA TESTING- RADIOGRAPHY (X-RAY -FIXED PORTABLE, MOBILE)	Dental Intraoral Periapical Radiograph (IOPA), Orthopantomogram (OPG) & Dental Hand Held	Operating Potential	AERB/RF-MED/SC-3 (Rev. 2)